



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 804M0201  
PRODUCT NAME: V-GRAD 80%  
REFERENCE: V-GRD-80  
EXPIRY DATE: 31.10.2019

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,41
OSMOLALITY	310 - 340 mOsm/kg	324
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED