

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 804M3201  
PRODUCT NAME: V-OIL 100 G  
REFERENCE: V-OIL-G100  
EXPIRY DATE: 31.10.2019

TEST	SPECIFICATION	RESULT
Peroxide value (POV) determination	< 0,1 mEq/kG	PASSED
ENDOTOXIN	PASSED	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
Blastocyst Cell Count	PASSED	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED

