



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 805M0301
PRODUCT NAME: V-GRAD 40%
REFERENCE: V-GRD-40
EXPIRY DATE: 30.11.2019

| TEST | SPECIFICATION | RESULT |
|---------------------|-------------------|--------|
| pH | 7,20-7,60 | 7,41 |
| OSMOLALITY | 310 - 340 mOsm/kg | 320 |
| BACTERIAL SCREEN | PASSED | PASSED |
| FUNGAL SCREEN | PASSED | PASSED |
| ENDOTOXIN | < 0,5 EU/mL | PASSED |
| SPERM SURVIVAL TEST | ≥ 80 % | PASSED |