

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 805M1401  
PRODUCT NAME: V-ONESTEP Culture Medium  
REFERENCE: V-OSM-20  
EXPIRY DATE: 22.08.2018

TEST	SPECIFICATION	RESULT
pH	7,25-7,40	7,31
OSMOLALITY	257-273 mOsm/kg	261
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED

**VITROMED GmbH**

Wildenbruchstr. 15  
07745 Jena - Germany  
Tel +49 3641539 -1976, Fax -1977  
office@vitromed.com  
www.vitromed.com

