

CERTIFICATE OF ANALYSIS

LOT NUMBER: 805M3401

PRODUCT NAME: V-PVP 7%

REFERENCE: V-PVP-7

EXPIRY DATE: 31.05.2019

TEST	SPECIFICATION	RESULT
pH	7,25-7,45	7,26
OSMOLALITY	300-330 mOsm/kg	309
ENDOTOXIN	< 0,5 EU/mL	< 0,05
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED
FUNGAL SCREEN	PASSED	PASSED
BACTERIAL SCREEN	PASSED	PASSED