



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 806M0501
PRODUCT NAME: V-SPERM WASH
REFERENCE: V-SWM-100
EXPIRY DATE: 30.06.2019

TEST	SPECIFICATION	RESULT
pH	7,00-7,40	7,31
OSMOLALITY	257-273 mOsm/kg	261
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED