



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 806M3401
PRODUCT NAME: V-PVP 7%
REFERENCE: V-PVP-7
EXPIRY DATE: 30.06.2019

TEST	SPECIFICATION	RESULT
pH	7,25-7,45	7,31
OSMOLALITY	300-330 mOsm/kg	306
ENDOTOXIN	< 0,5 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED
FUNGAL SCREEN	PASSED	PASSED
BACTERIAL SCREEN	PASSED	PASSED