

CERTIFICATE OF ANALYSIS

LOT NUMBER: 806M3601

PRODUCT NAME: V-HYLASE

REFERENCE: V-HYA

EXPIRY DATE: 30.06.2019

TEST	SPECIFICATION	RESULT
pH	7,20-7,40	7,27
OSMOLALITY	257-273 mOsm/kg	269
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,00 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED