



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 807M0201
PRODUCT NAME: V-GRAD 80%
REFERENCE: V-GRD-80
EXPIRY DATE: 31.01.2020

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,40
OSMOLALITY	310-330 mOsm/kg	314
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED