



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 807M1301  
PRODUCT NAME: V-ONESTEP Culture Medium  
REFERENCE: V-OSM-10  
EXPIRY DATE: 07.10.2018

TEST	SPECIFICATION	RESULT
pH	7,25-7,40	7,35
OSMOLALITY	257-273 mOsm/kg	263
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED