



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 808M1302
PRODUCT NAME: V-ONESTEP Culture Medium
REFERENCE: V-OSM-10
EXPIRY DATE: 26.11.2018

TEST	SPECIFICATION	RESULT
pH	7,25-7,40	7,33
OSMOLALITY	257-273 mOsm/kg	257
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED
Blastocyst Cell Count	PASSED	PASSED