



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 808M3002  
PRODUCT NAME: V-OIL 100 P  
REFERENCE: V-OIL-P100  
EXPIRY DATE: 29.02.2020

TEST	SPECIFICATION	RESULT
Peroxide value (POV) determination	< 0,1 mEq/kg	PASSED
ENDOTOXIN	< 0,1 EU/ mL	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
Blastocyst Cell Count	PASSED	PASSED
Time Lapse Mouse Embryo Assay	≥ 80 %	PASSED

**VITROMED GmbH**  
Wildenbruchstr. 15  
07745 Jena - Germany  
Tel +49 3641539 -1976, Fax -1977  
office@vitromed.com  
www.vitromed.com

i.A.