



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 809M0601  
PRODUCT NAME: V-SPERM FREEZE  
REFERENCE: V-SFM-12  
EXPIRY DATE: 31.03.2020

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,50
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED

**VITROMED GmbH**  
Wildenbruchstr. 15  
07745 Jena - Germany  
Tel +49 3641539 -1976, Fax -1977  
office@vitromed.com  
www.vitromed.com