



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 809M1302  
PRODUCT NAME: V-ONESTEP Culture Medium  
REFERENCE: V-OSM-10  
EXPIRY DATE: 24.12.2018

TEST	SPECIFICATION	RESULT
pH	7,25-7,40	7,33
OSMOLALITY	257-273 mOsm/kg	262
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED