



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 810M0401
PRODUCT NAME: V-GRAD Kit
REFERENCE: V-GRD-KIT
EXPIRY DATE: 30.04.2020

TEST	SPECIFICATION	RESULT
pH (V-GRAD 40)	7,20-7,60	7,34
pH (V-GRAD 80)	7,20-7,60	7,38
OSMOLALITY (V-GRAD 40)	310 - 340 mOsm/kg	319
OSMOLALITY (V-GRAD 80)	310-330 mOsm/kg	316
ENDOTOXIN	< 0,5 EU/mL	PASSED
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED