



VITROMED
Germany

VITROMED GmbH
Raiffeisenstr 15a
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 911VM2302
PRODUCT NAME: V-HEPES PLUS 100
REFERENCE: V-HTF-P1
EXPIRY DATE: 20.11.2020

TEST	SPECIFICATION	RESULT
pH	7.00-7.40	7.31
OSMOLALITY	257-273 mOsm/kg	270
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED
Color	PASSED	PASSED