



VITROMED  
Germany

VITROMED GmbH  
Raiffeisenstraße 15a  
40764 Langenfeld  
GERMANY

Tel +49 2173-20041-30

www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 912VM3503  
PRODUCT NAME: V-PVP 10%  
REFERENCE: V-PVP-10  
EXPIRY DATE: 17-Dec-2020

TEST	SPECIFICATION	RESULT
pH	7,25 – 7,45	7,43
OSMOLALITY	300 – 330 mOsm/kg	314
BACTERIAL SCREEN	< 0,5 EU/ml	PASSED
FUNGAL SCREEN	≥ 80 %	PASSED
ENDOTOXIN	PASSED	PASSED
MEA 1-cell to expanded blastocyst 96h	PASSED	PASSED